

**DFM / UWMF Clinics: Medication Refrigerator/Freezer Temperature Log**

Location: \_\_\_\_\_

Month/Year: \_\_\_\_\_

***Appropriate Refrigerator Range = 36° F to 46° F***

***Appropriate Freezer Range = 5° F or LOWER***

day	REFRIGERATOR								FREEZER				Doors closed at end of day
	AM				PM				AM		PM		
	Temp (F°)	Min	Max	Initials / Comments	Temp (F°)	Min	Max	Initials / Comments	Temp (F°)	Initials	Temp (F°)	Initials / Comments	
1													
2													
3													
4													
5													
6													
7													
8													
9													
10													
11													
12													
13													
14													
15													
16													
17													
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													

In the event that the refrigerator and/or freezer temperature (**including min/max temperature values for refrigerator**) are not within the accepted range, staff shall **IMMEDIATELY NOTIFY the clinic manager or nursing leader** who will seek consultation to determine if medications and antigens remain viable.